

# LUNG VENTILATION STUDY

## (Xe-133 Gas)

### Overview

- The Lung Ventilation Study demonstrates the distribution of ventilation, air space, and air trapping within the lungs in the posterior projection, but images in other projections can be acquired secondarily.

### Indications

- Diagnosis of pulmonary embolism, particularly when helical CT is contraindicated because of renal insufficiency or a history of a bona fide contrast reaction (1-4).
- Evaluation of regional ventilation (5,6).

### Examination Time

- Initial POST images: 20 minutes.
- Optional reventilation images in other projections: 20 minutes each additional projection (7-9).

### Patient Preparation

- Rehearse the patient through the breathing maneuvers required for image acquisition.

### Equipment & Energy Windows

- Gamma camera: Large field of view.
- Collimator: Low energy, high resolution, parallel hole.
- Energy window: 20% window centered at 80 keV.
- Gas dispenser with return trap and 3 way valve.

### Radiopharmaceutical, Dose, & Technique of Administration

- Radiopharmaceutical (10): Xe-133 gas.
- Dose: 20 mCi (740 MBq).
- Technique of administration: Xenon delivery system with trap:
  1. Fit the patient with a tightly fitting mask or a mouth piece and nose clamp.
  2. Attach the xenon delivery system for injection of Xe-133 gas and collection of exhaled Xe-133 gas.
  3. Set the valves so the patient is breathing from and into the xenon system, i.e. a closed system.


## Patient Position & Imaging Field

- Patient position: Sitting (supine if unable to sit).
- Imaging field: Entire lungs; the Xe-133 dose may be used as a transmission source to ensure that the lungs are all within the field of view.

## Acquisition Protocol (4)

- Be sure a new external filter is in the gas delivery system.
- Acquire images in the POST projection.
- Acquire a single breath (ventilation) analog image:
  1. Instruct the patient to take a deep breath as the Xe-133 gas bolus is injected into the delivery system and then hold the breath as long as possible.
  2. Acquire a 100 K count image.
  3. If the patient starts to breathe before 100 K counts are acquired, immediately terminate the acquisition.
- Equilibrate the concentration of Xe-133 gas within the patient's lungs:
  1. Have the patient breathe normally for 3 minutes.
- Acquire an equilibrium (airspace) analog image:
  1. Acquire an approximately 300 K count image.
- Acquire a series of washout (airway obstruction) analog images:
  1. Change the system valve so that the patient breathes room air in and exhales Xe-133 into the system trap.
  2. Beginning immediately, acquire sequential 30 second analog images until the Xe-133 gas is gone as judged from the persistence scope. Acquire a minimum of 4 images.
- Close the xenon delivery system and remove the mask from the patient's face.

## Protocol Summary Diagram



## Data Processing

- None.

## Optional Maneuvers

- Reventilation studies (7,8): If after seeing the initial ventilation study and the perfusion study, the nuclear medicine physician determines that a reventilation study is needed in another projection, proceed as follows:
  1. Reposition the patient in the selected projection; emphasize to the patient the importance of not moving throughout the entire image acquisition sequence.
  2. Acquire a repeat perfusion image using the Tc-99m energy window.
  3. Acquire a “Tc-99m scatter image” using the Xe-133 energy window. Acquire this image for the same time that the single breath image requires, approximately 20 seconds.
  4. Repeat the ventilation acquisition outlined above without moving the patient.
- Alternative radiopharmaceuticals: In some countries and localities Kr-81m gas (11) and Xe-127 gas (12) are available and may be substituted for Xe-133. Some protocol changes are necessary.

## Principle Radiation Emission Data - Xe-133 (13)

- Physical half-life = 5.25 days.

Radiation	Mean % per disintegration	Mean energy (keV)
Beta-2	99.3	100.5
Ce-K-2	52.0	45.0
Ce-L-2	8.5	75.3
Ce-M-2	2.3	79.8
Gamma-2	37.1	81.0
K alpha 2 x-ray	13.3	30.6
K alpha 1 x-ray	24.6	31.0
K beta x-rays	8.8	35.0

## Dosimetry - Xe-133 Gas (14-16)

Organ	rads/20 mCi	mGy/740 MBq
Lungs	0.17	1.7
Whole body	0.002	0.02
Brain	0.001	0.01

## Dosimetry - Kr-81m Gas (14)

Organ	rads/mCi-min	mGy/MBq-min
Tracheal mucosa (surface)	0.46	0.12
Lungs	0.0025	0.0067
Whole body	0.000067	0.000018

## References

1. Biello DR, Mattar AG, McKnight RC, et al: Ventilation perfusion studies in suspected pulmonary embolism. *Am J Roentgenol* 133:1033-1037, 1979.
2. Alderson PO, Biello DR, Gottschalk A: Tc-99m DTPA aerosol and radioactive gases compared as adjuncts to perfusion scintigraphy in patients with suspected pulmonary embolism. *Radiology* 153:515-521, 1984.
3. Coche E, Verschuren F, Keyeux A, et al: Diagnosis of acute pulmonary embolism in outpatients: Comparison of thin-collimation multi-detector row spiral cT and planar

- ventilation-perfusion scintigraphy. Radiology 229:757-765, 2003.
4. Klingensmith WC, Holt SA: Lung scan interpretation: A user-friendly, physiologic approach. J Nucl Med 33:1417-1422, 1992.
  5. Alderson PO, Secker-Walker RH, Forrest JV: Detection of obstructive pulmonary disease: Relative sensitivity of ventilation-perfusion studies and chest radiography. Radiology 11:643-648, 1974.
  6. Ali MK, Mountain CF, Ewer MS, et al: Predicting loss of pulmonary function after pulmonary resection for bronchogenic carcinoma. Chest 77:337-342, 1980.
  7. Jacobstein JG: Xe-133 ventilation scanning immediately following the Tc-99m perfusion scan. J Nucl Med 15:964-968, 1974.
  8. Kipper MS, Alazraki N: The feasibility of performing Xe-133 ventilation imaging following the perfusion study. Radiology 144:581-586, 1982.
  9. Stein MG, Waxman AD, Ramanna L, et al: Postperfusion xenon-133 ventilation scintigraphy with a narrow window. Am J Roentgenol 145:511-515, 1985.
  10. Ramanna L, Alderson PO, Waxman AD, et al: Regional comparison of technetium-99m DTPA aerosol and radioactive gas ventilation (xenon and krypton) studies in patients with suspected pulmonary embolism. J Nucl Med 27:1391-1396, 1986.
  11. Parker AJ, Coleman RE, Siegel BA, et al: Procedure guideline for lung scintigraphy: 1.0. J Nucl Med 37:1906-1910, 1996.
  12. Goddard BA, Ackery DM: Xenon-133, Xe-127 and Xe-125 for lung function investigations: A dosimetric comparison. J Nucl Med 16:780-786, 1975.
  13. 54-Xe-133: In MIRD: Radionuclide Data and Decay Schemes, DA Weber, KF Eckerman, AT Dillman, JC Ryman, eds, Society of Nuclear Medicine, New York, 1989, pp 246-247.
  14. Atkins HL, Robertson JS, Croft BY, et al: MIRD Dose Estimate Report No 9: Estimates of radiation absorbed doses from radioxenons in lung imaging. J Nucl Med 21:459-465, 1980.
  15. Atkins HL, Robertson JS, Akabani G: MIRD dose estimate report no. 17: Radiation absorbed dose estimates from inhaled krypton-81m gas in lung imaging. J Nucl Med 34:1382-1384, 1993.
  16. Prohovnik I, Metz CD, Atkins HL: Radiation exposure to human trachea from xenon-133 procedures. J Nucl Med 36:1458-1461, 1995.

#### Normal Findings

- > Stein MG, Waxman AD, Ramanna L, et al: Postperfusion xenon-133 ventilation scintigraphy with a narrow window. Am J Roentgenol 145:511-515, 1985.
- > Morrell NW, Roberts CM, Jones BE, et al: The anatomy of radioisotope lung scanning. J Nucl Med 33:676-683, 1992.

[Return to Table of Content](#)